

APPENDIX N.2 – ATTACHMENT 1
EUROFINS CS 2014 PE SAMPLE RESULTS



QUALITY ASSURANCE TECHNICAL SUPPORT LABORATORY PES SCORING EVALUATION REPORT

PES:	SRS1830	EPA Sample No.:	SRS1830	Report Date:	07/11/2014
Lab Name:	Calscience Environmental Laboratories, Inc.	Case No.:	PV Shelf	Lab Code:	Calscience
Contract:	14-05-1480	Matrix:	Sediment	SAS/Client No.:	NR
SDG No.:	NR	Date Received:	05/20/2014	Lab Sample ID:	14-05-1480-1-A
Lab File ID:	NR	Sample Wt./Vol. (g/mL):	NR	Date Extracted:	06/04/2014
Date Analyzed:	06/10/2014	Decanted:	No	Level:	NA
% Moisture:	0.0%	GPC Cleanup:	NR	Conc. Extract Vol. (µL):	NR
Inject. Vol. (µL):	NR	Extraction Type:	NR	pH:	NA
Dilution Factor:	5.0	GPC Factor:	NR	Units:	µg/Kg
Analysis Method:	EPA Method 8270C	Sulfur Cleanup:	NR		
Scoring Method:	Manual				
Comments:	Scored by CB&I Personnel				

CAS No.	Analyte	Laboratory Results		PES Evaluation	
		Concentration	Q		
72-54-8	4,4'-DDD	1900		Pass	Within Limits
72-55-9	4,4'-DDE	1400		Pass	Within Limits
50-29-3	4,4'-DDT	2000		Pass	Within Limits
53-19-0	2,4'-DDD	1200		Pass	Within Limits
3424-82-6	2,4'-DDE	1300		Pass	Within Limits
789-02-6	2,4'-DDT	1200		Pass	Within Limits
****	Total DDT	9000		Pass	Within Limits
****	END All Analytes	*****	****	*****	*****

NL = No Limit
 NR = Not Reported
 NE = Not Evaluated
 NA = Not Applicable



QUALITY ASSURANCE TECHNICAL SUPPORT LABORATORY PES SCORING EVALUATION REPORT

PES:	SRS1830	EPA Sample No.:	SRS1830	Report Date:	07/11/2014
Lab Name:	Calscience Environmental Laboratories, Inc.			Lab Code:	Calscience
Contract:	14-05-1480	Case No.:	PV Shelf	SAS/Client No.:	NR
SDG No.:	NR	Matrix:	Sediment	Lab Sample ID:	14-05-1480-1-A
Lab File ID:	NR	Date Received:	05/20/2014	Date Extracted:	06/04/2014
Date Analyzed:	06/10/2014	Sample Wt./Vol. (g/mL):	NR	Level:	NA
% Moisture:	0.0%	Decanted:	No	Conc. Extract Vol. (µL):	NR
Inject. Vol. (µL):	NR	GPC Cleanup:	NR	pH:	NA
Dilution Factor:	1.0	Extraction Type:	NR	Units:	µg/Kg
Analysis Method:	EPA Method 8270C	GPC Factor:	NR		
Scoring Method:	Manual	Sulfur Cleanup:	NR		
Comments:	Scored by CB&I Personnel				

PCB Number	Laboratory Results		PES Evaluation	
	Concentration	Q		
008	0.37		Pass	Within Limits
018	0.83		Pass	Within Limits
028	<0.20		NE	NE
031	<0.20		NE	NE
033	<0.20		NE	NE
044	1.8		Pass	Within Limits
049	1.2		Pass	Within Limits
052	2.5		Pass	Within Limits
066	1.1		Pass	Within Limits
070/074	1.9		Pass	Within Limits
087/097	<0.20		NE	NE
095/098	3.8		Pass	Within Limits
099	1.5		Pass	Within Limits
101	4.2		Pass	Within Limits
105	<0.20		NE	NE
110	3.6		Pass	Within Limits
118	2.6		Pass	Within Limits
128	0.81		Pass	Within Limits
132/153/168	14		Pass	Within Limits
129/138/160/163	9.9		Pass	Within Limits
141	2.5		Pass	Within Limits
147/149	6.4		Pass	Warning Low
135/151/154	2.5		Pass	Warning Low
156/157	0.50		Pass	Within Limits
158	9.9		NE	NE
170	3.2		Pass	Within Limits
174	4.6		Pass	Within Limits
177	1.8		Pass	Within Limits
180/193	8.6		Pass	Within Limits
183/185	1.9		Pass	Warning Low



QUALITY ASSURANCE TECHNICAL SUPPORT LABORATORY PES SCORING EVALUATION REPORT

PCB Number	Laboratory Results		PES Evaluation	
	Concentration	Q		
187	4.1		Pass	Within Limits
194	2.0		Pass	Within Limits
195	0.94		Pass	Within Limits
201	0.28		Pass	Within Limits
203	1.7		Pass	Within Limits
206	0.62		Pass	Within Limits
Total PCBs	NR		NE	NE
END All Analytes	*****	****	*****	*****

NL = No Limit
NR = Not Reported
NE = Not Evaluated
NA = Not Applicable



QUALITY ASSURANCE TECHNICAL SUPPORT LABORATORY PES SCORING EVALUATION REPORT

PES: SRS1830 **EPA Sample No.:** SRS1830 **Report Date:** 07/11/2014
Lab Name: Calscience Environmental Laboratories, Inc. **Lab Code:** Calscience
Contract: 14-05-1480 **Case No.:** PV Shelf **SAS/Client No.:** NR
SDG No.: NR **Matrix:** Sediment **Lab Sample ID:** 14-05-1480-1-A
Lab File ID: NR **Date Received:** 05/20/2014 **Date Extracted:** NA
Date Analyzed: 06/06/2014 **Sample Wt./Vol. (g/mL):** NR **Level:** NA
% Moisture: 0.0% **Decanted:** NA **Conc. Extract Vol. (μL):** NR
Inject. Vol. (μL): NR **GPC Cleanup:** NA **pH:** NA
Dilution Factor: NA **Extraction Type:** NA **Units:** %
Analysis Method: EPA Method 9060A **GPC Factor:** NA
Scoring Method: Manual **Sulfur Cleanup:** NA
Comments: Scored by CB&I Personnel

CAS No.	Analyte	Laboratory Results		PES Evaluation	
		Concentration	Q		
****	Total Organic Carbon	2.2		Pass	Within Limits
****	END All Analytes	*****	****	*****	*****

NL = No Limit
NR = Not Reported
NE = Not Evaluated
NA = Not Applicable



WORK ORDER NUMBER: 14-05-1480

The difference is service



AIR | SOIL | WATER | MARINE CHEMISTRY

Analytical Report For

Client: Gilbane

Client Project Name: PALOS VERDE SHELF

Attention: Kristen Carlyon
2730 Shadelands Drive, Suite 100
Walnut Creek, CA 94598-2540

Danielle Gonsman

Approved for release on 06/25/2014 by:
Danielle Gonsman
Project Manager

ResultLink ▶

Email your PM ▶



Eurofins Calscience, Inc. (Calscience) certifies that the test results provided in this report meet all NELAC requirements for parameters for which accreditation is required or available. Any exceptions to NELAC requirements are noted in the case narrative. The original report of subcontracted analyses, if any, is attached to this report. The results in this report are limited to the sample(s) tested and any reproduction thereof must be made in its entirety. The client or recipient of this report is specifically prohibited from making material changes to said report and, to the extent that such changes are made, Calscience is not responsible, legally or otherwise. The client or recipient agrees to indemnify Calscience for any defense to any litigation which may arise.

Contents

Client Project Name: PALOS VERDE SHELF
Work Order Number: 14-05-1480

1	Work Order Narrative.	3
2	Sample Summary.	4
3	Client Sample Data.	5
	3.1 EPA 9060A Total Organic Carbon (Solid).	5
	3.2 EPA 8270C SIM Pesticides (Solid).	6
	3.3 EPA 8270C SIM PCB Congeners (Solid).	8
4	Quality Control Sample Data.	12
	4.1 LCS/LCSD.	12
5	Glossary of Terms and Qualifiers.	15
6	Chain-of-Custody/Sample Receipt Form.	16

Work Order Narrative

Work Order: 14-05-1480

Page 1 of 1

Condition Upon Receipt:

Samples were received under Chain-of-Custody (COC) on 05/20/14. They were assigned to Work Order 14-05-1480.

Unless otherwise noted on the Sample Receiving forms all samples were received in good condition and within the recommended EPA temperature criteria for the methods noted on the COC. The COC and Sample Receiving Documents are integral elements of the analytical report and are presented at the back of the report.

Holding Times:

All samples were analyzed within prescribed holding times (HT) and/or in accordance with the Calscience Sample Acceptance Policy unless otherwise noted in the analytical report and/or comprehensive case narrative, if required.

Any parameter identified in 40CFR Part 136.3 Table II that is designated as "analyze immediately" with a holding time of ≤ 15 minutes (40CFR-136.3 Table II, footnote 4), is considered a "field" test and the reported results will be qualified as being received outside of the stated holding time unless received at the laboratory within 15 minutes of the collection time.

Quality Control:

All quality control parameters (QC) were within established control limits except where noted in the QC summary forms or described further within this report.

Additional Comments:

Air - Sorbent-extracted air methods (EPA TO-4A, EPA TO-10, EPA TO-13A, EPA TO-17): Analytical results are converted from mass/sample basis to mass/volume basis using client-supplied air volumes.

New York NELAP air certification does not certify for all reported methods and analytes, reference the accredited items here: http://www.calscience.com/PDF/New_York.pdf

Solid - Unless otherwise indicated, solid sample data is reported on a wet weight basis, not corrected for % moisture. All QC results are always reported on a wet weight basis.

Subcontractor Information:

Unless otherwise noted below (or on the subcontract form), no samples were subcontracted.



Calscience

Sample Summary

Client:	Gilbane	Work Order:	14-05-1480
	2730 Shadelands Drive, Suite 100	Project Name:	PALOS VERDE SHELF
	Walnut Creek, CA 94598-2540	PO Number:	
		Date/Time Received:	05/20/14 10:30
		Number of Containers:	1

Attn: Kristen Carlyon

Sample Identification	Lab Number	Collection Date and Time	Number of Containers	Matrix
SRS1830	14-05-1480-1	05/19/14 00:00	1	Sediment


Return to Contents



Calscience

Analytical Report

Gilbane	Date Received:	05/20/14
2730 Shadelands Drive, Suite 100	Work Order:	14-05-1480
Walnut Creek, CA 94598-2540	Preparation:	N/A
	Method:	EPA 9060A
	Units:	mg/kg
Project: PALOS VERDE SHELF		Page 1 of 1

Client Sample Number	Lab Sample Number	Date/Time Collected	Matrix	Instrument	Date Prepared	Date/Time Analyzed	QC Batch ID
SRS1830	14-05-1480-1-A	05/19/14 00:00	Sediment	TOC 5	06/06/14	06/06/14 14:40	E0606TOCL1

Comment(s): - Results are reported on a dry weight basis.
 - Results were evaluated to the MDL (DL), concentrations $\geq$ to the MDL (DL) but $<$ RL (LOQ), if found, are qualified with a "J" flag.

<u>Parameter</u>	<u>Result</u>	<u>RL</u>	<u>MDL</u>	<u>DF</u>	<u>Qualifiers</u>
Carbon, Total Organic	22000	500	120	1.00	

Method Blank	099-06-013-1057	N/A	Solid	TOC 5	06/06/14	06/06/14 14:40	E0606TOCL1
---------------------	------------------------	------------	--------------	--------------	-----------------	-----------------------	-------------------

Comment(s): - Results were evaluated to the MDL (DL), concentrations $\geq$ to the MDL (DL) but $<$ RL (LOQ), if found, are qualified with a "J" flag.

<u>Parameter</u>	<u>Result</u>	<u>RL</u>	<u>MDL</u>	<u>DF</u>	<u>Qualifiers</u>
Carbon, Total Organic	ND	500	120	1.00	

Return to Contents

RL: Reporting Limit. DF: Dilution Factor. MDL: Method Detection Limit.



Calscience

Analytical Report

Gilbane
2730 Shadelands Drive, Suite 100
Walnut Creek, CA 94598-2540

Date Received: 05/20/14
Work Order: 14-05-1480
Preparation: EPA 3545
Method: EPA 8270C PEST-SIM
Units: ug/kg

Project: PALOS VERDE SHELF

Page 1 of 2

Client Sample Number	Lab Sample Number	Date/Time Collected	Matrix	Instrument	Date Prepared	Date/Time Analyzed	QC Batch ID
SRS1830	14-05-1480-1-A	05/19/14 00:00	Sediment	GC/MS HHH	06/04/14	06/10/14 21:38	140604L31

Comment(s): - Results are reported on a dry weight basis.
- Results were evaluated to the MDL (DL), concentrations $\geq$ to the MDL (DL) but $<$ RL (LOQ), if found, are qualified with a "J" flag.

Parameter	Result	RL	MDL	DF	Qualifiers
4,4'-DDMU	ND	5.0	1.4	1.00	
4,4'-DDNU	ND	5.0	1.5	1.00	

Surrogate	Rec. (%)	Control Limits	Qualifiers
2,4,5,6-Tetrachloro-m-Xylene	83	50-125	

SRS1830	14-05-1480-1-A	05/19/14 00:00	Sediment	GC/MS HHH	06/04/14	06/10/14 15:21	140604L31
---------	----------------	----------------	----------	-----------	----------	----------------	-----------

Comment(s): - Results are reported on a dry weight basis.
- Results were evaluated to the MDL (DL), concentrations $\geq$ to the MDL (DL) but $<$ RL (LOQ), if found, are qualified with a "J" flag.

Parameter	Result	RL	MDL	DF	Qualifiers
2,4'-DDD	1200	25	12	5.00	
2,4'-DDE	1300	25	7.4	5.00	
2,4'-DDT	1200	25	14	5.00	
4,4'-DDD	1900	25	15	5.00	
4,4'-DDE	1400	25	9.9	5.00	
4,4'-DDT	2000	25	16	5.00	

Surrogate	Rec. (%)	Control Limits	Qualifiers
2,4,5,6-Tetrachloro-m-Xylene	90	50-125	

RL: Reporting Limit. DF: Dilution Factor. MDL: Method Detection Limit.

Analytical Report

Gilbane
2730 Shadelands Drive, Suite 100
Walnut Creek, CA 94598-2540

Date Received: 05/20/14
Work Order: 14-05-1480
Preparation: EPA 3545
Method: EPA 8270C PEST-SIM
Units: ug/kg

Project: PALOS VERDE SHELF

Page 2 of 2

Client Sample Number	Lab Sample Number	Date/Time Collected	Matrix	Instrument	Date Prepared	Date/Time Analyzed	QC Batch ID
Method Blank	099-13-013-73	N/A	Solid	GC/MS HHH	06/04/14	06/10/14 13:36	140604L31

Comment(s): - Results were evaluated to the MDL (DL), concentrations $\geq$ to the MDL (DL) but $<$ RL (LOQ), if found, are qualified with a "J" flag.

<u>Parameter</u>	<u>Result</u>	<u>RL</u>	<u>MDL</u>	<u>DF</u>	<u>Qualifiers</u>
2,4'-DDD	ND	5.0	2.3	1.00	
2,4'-DDE	ND	5.0	1.5	1.00	
2,4'-DDT	ND	5.0	2.9	1.00	
4,4'-DDD	ND	5.0	3.0	1.00	
4,4'-DDE	ND	5.0	2.0	1.00	
4,4'-DDMU	ND	5.0	1.4	1.00	
4,4'-DDNU	ND	5.0	1.5	1.00	
4,4'-DDT	ND	5.0	3.2	1.00	

<u>Surrogate</u>	<u>Rec. (%)</u>	<u>Control Limits</u>	<u>Qualifiers</u>
2,4,5,6-Tetrachloro-m-Xylene	83	50-125	

Return to Contents

RL: Reporting Limit. DF: Dilution Factor. MDL: Method Detection Limit.



Calscience

Analytical Report

Gilbane
2730 Shadelands Drive, Suite 100
Walnut Creek, CA 94598-2540

Date Received: 05/20/14
Work Order: 14-05-1480
Preparation: EPA 3545
Method: EPA 8270C SIM PCB Congeners
Units: ug/kg

Project: PALOS VERDE SHELF

Page 1 of 4

Client Sample Number	Lab Sample Number	Date/Time Collected	Matrix	Instrument	Date Prepared	Date/Time Analyzed	QC Batch ID
SRS1830	14-05-1480-1-A	05/19/14 00:00	Sediment	GC/MS HHH	06/04/14	06/10/14 22:10	140604L30

Comment(s): - Results are reported on a dry weight basis.

- Results were evaluated to the MDL (DL), concentrations $\geq$ to the MDL (DL) but $<$ RL (LOQ), if found, are qualified with a "J" flag.

Parameter	Result	RL	MDL	DF	Qualifiers
PCB003	ND	0.20	0.047	1.00	
PCB005/008	0.37	0.40	0.047	1.00	J
PCB015	ND	0.20	0.023	1.00	
PCB018	0.83	0.20	0.039	1.00	
PCB027	ND	0.20	0.028	1.00	
PCB028	ND	0.20	0.055	1.00	
PCB029	ND	0.20	0.035	1.00	
PCB031	ND	0.20	0.029	1.00	
PCB033	ND	0.20	0.035	1.00	
PCB037	ND	0.20	0.035	1.00	
PCB044	1.8	0.20	0.092	1.00	
PCB049	1.2	0.20	0.086	1.00	
PCB052	2.5	0.20	0.051	1.00	
PCB056	ND	0.20	0.048	1.00	
PCB060	ND	0.20	0.063	1.00	
PCB066	1.1	0.20	0.075	1.00	
PCB070	1.3	0.20	0.048	1.00	
PCB074	0.63	0.20	0.046	1.00	
PCB077	0.65	0.20	0.085	1.00	
PCB081	ND	0.20	0.064	1.00	
PCB087	ND	0.20	0.041	1.00	
PCB095	3.8	0.20	0.050	1.00	
PCB097	ND	0.20	0.043	1.00	
PCB099	1.5	0.20	0.054	1.00	
PCB101	4.2	0.20	0.051	1.00	
PCB105	ND	0.20	0.042	1.00	
PCB110	3.6	0.20	0.046	1.00	
PCB114	ND	0.20	0.036	1.00	
PCB118	2.6	0.20	0.059	1.00	
PCB119	ND	0.20	0.046	1.00	
PCB123	ND	0.20	0.047	1.00	
PCB126	ND	0.20	0.034	1.00	
PCB128	0.81	0.20	0.039	1.00	

RL: Reporting Limit. DF: Dilution Factor. MDL: Method Detection Limit.



Calscience

Analytical Report

Gilbane
2730 Shadelands Drive, Suite 100
Walnut Creek, CA 94598-2540

Date Received: 05/20/14
Work Order: 14-05-1480
Preparation: EPA 3545
Method: EPA 8270C SIM PCB Congeners
Units: ug/kg

Project: PALOS VERDE SHELF

Page 2 of 4

<u>Parameter</u>	<u>Result</u>	<u>RL</u>	<u>MDL</u>	<u>DF</u>	<u>Qualifiers</u>
PCB132/153	14	0.40	0.067	1.00	
PCB137	ND	0.20	0.048	1.00	
PCB138/158	9.9	0.40	0.075	1.00	
PCB141	2.5	0.20	0.098	1.00	
PCB149	6.4	0.20	0.048	1.00	
PCB151	2.5	0.20	0.062	1.00	
PCB156	0.50	0.20	0.066	1.00	
PCB157	ND	0.20	0.051	1.00	
PCB167	ND	0.20	0.042	1.00	
PCB168	ND	0.20	0.045	1.00	
PCB169	ND	0.20	0.033	1.00	
PCB170	3.2	0.20	0.050	1.00	
PCB174	4.6	0.20	0.039	1.00	
PCB177	1.8	0.20	0.040	1.00	
PCB180	8.6	0.20	0.030	1.00	
PCB183	1.9	0.20	0.032	1.00	
PCB184	ND	0.20	0.041	1.00	
PCB187	4.1	0.20	0.039	1.00	
PCB189	ND	0.20	0.025	1.00	
PCB194	2.0	0.20	0.041	1.00	
PCB195	0.94	0.20	0.032	1.00	
PCB200	ND	0.20	0.059	1.00	
PCB201	0.28	0.20	0.044	1.00	
PCB203	1.7	0.20	0.035	1.00	
PCB206	0.62	0.20	0.045	1.00	
PCB209	0.21	0.20	0.067	1.00	

<u>Surrogate</u>	<u>Rec. (%)</u>	<u>Control Limits</u>	<u>Qualifiers</u>
2-Fluorobiphenyl	77	50-150	
p-Terphenyl-d14	73	50-150	

Return to Contents

RL: Reporting Limit. DF: Dilution Factor. MDL: Method Detection Limit.



Calscience

Analytical Report

Gilbane
2730 Shadelands Drive, Suite 100
Walnut Creek, CA 94598-2540

Date Received: 05/20/14
Work Order: 14-05-1480
Preparation: EPA 3545
Method: EPA 8270C SIM PCB Congeners
Units: ug/kg

Project: PALOS VERDE SHELF

Page 3 of 4

Client Sample Number	Lab Sample Number	Date/Time Collected	Matrix	Instrument	Date Prepared	Date/Time Analyzed	QC Batch ID
Method Blank	099-16-418-9	N/A	Solid	GC/MS HHH	06/04/14	06/16/14 20:07	140604L30

Comment(s): - Results were evaluated to the MDL (DL), concentrations $\geq$ to the MDL (DL) but $<$ RL (LOQ), if found, are qualified with a "J" flag.

Parameter	Result	RL	MDL	DF	Qualifiers
PCB003	ND	0.20	0.047	1.00	
PCB005/008	ND	0.40	0.047	1.00	
PCB015	ND	0.20	0.023	1.00	
PCB018	ND	0.20	0.039	1.00	
PCB027	ND	0.20	0.028	1.00	
PCB028	ND	0.20	0.055	1.00	
PCB029	ND	0.20	0.035	1.00	
PCB031	ND	0.20	0.029	1.00	
PCB033	ND	0.20	0.035	1.00	
PCB037	ND	0.20	0.035	1.00	
PCB044	ND	0.20	0.092	1.00	
PCB049	ND	0.20	0.086	1.00	
PCB052	ND	0.20	0.051	1.00	
PCB056	ND	0.20	0.048	1.00	
PCB060	ND	0.20	0.063	1.00	
PCB066	ND	0.20	0.075	1.00	
PCB070	ND	0.20	0.048	1.00	
PCB074	ND	0.20	0.046	1.00	
PCB077	ND	0.20	0.085	1.00	
PCB081	ND	0.20	0.064	1.00	
PCB087	ND	0.20	0.041	1.00	
PCB095	ND	0.20	0.050	1.00	
PCB097	ND	0.20	0.043	1.00	
PCB099	ND	0.20	0.054	1.00	
PCB101	ND	0.20	0.051	1.00	
PCB105	ND	0.20	0.042	1.00	
PCB110	ND	0.20	0.046	1.00	
PCB114	ND	0.20	0.036	1.00	
PCB118	ND	0.20	0.059	1.00	
PCB119	ND	0.20	0.046	1.00	
PCB123	ND	0.20	0.047	1.00	
PCB126	ND	0.20	0.034	1.00	
PCB128	ND	0.20	0.039	1.00	
PCB132/153	ND	0.40	0.067	1.00	

RL: Reporting Limit. DF: Dilution Factor. MDL: Method Detection Limit.



Calscience

Analytical Report

Gilbane
2730 Shadelands Drive, Suite 100
Walnut Creek, CA 94598-2540

Date Received: 05/20/14
Work Order: 14-05-1480
Preparation: EPA 3545
Method: EPA 8270C SIM PCB Congeners
Units: ug/kg

Project: PALOS VERDE SHELF

Page 4 of 4

<u>Parameter</u>	<u>Result</u>	<u>RL</u>	<u>MDL</u>	<u>DF</u>	<u>Qualifiers</u>
PCB137	ND	0.20	0.048	1.00	
PCB138/158	ND	0.40	0.075	1.00	
PCB141	ND	0.20	0.098	1.00	
PCB149	ND	0.20	0.048	1.00	
PCB151	ND	0.20	0.062	1.00	
PCB156	ND	0.20	0.066	1.00	
PCB157	ND	0.20	0.051	1.00	
PCB167	ND	0.20	0.042	1.00	
PCB168	ND	0.20	0.045	1.00	
PCB169	ND	0.20	0.033	1.00	
PCB170	ND	0.20	0.050	1.00	
PCB174	ND	0.20	0.039	1.00	
PCB177	ND	0.20	0.040	1.00	
PCB180	ND	0.20	0.030	1.00	
PCB183	ND	0.20	0.032	1.00	
PCB184	ND	0.20	0.041	1.00	
PCB187	ND	0.20	0.039	1.00	
PCB189	ND	0.20	0.025	1.00	
PCB194	ND	0.20	0.041	1.00	
PCB195	ND	0.20	0.032	1.00	
PCB200	ND	0.20	0.059	1.00	
PCB201	ND	0.20	0.044	1.00	
PCB203	ND	0.20	0.035	1.00	
PCB206	ND	0.20	0.045	1.00	
PCB209	ND	0.20	0.067	1.00	

<u>Surrogate</u>	<u>Rec. (%)</u>	<u>Control Limits</u>	<u>Qualifiers</u>
2-Fluorobiphenyl	89	50-150	
p-Terphenyl-d14	77	50-150	

Return to Contents

RL: Reporting Limit. DF: Dilution Factor. MDL: Method Detection Limit.



Calscience

Quality Control - LCS/LCSD

Gilbane
2730 Shadelands Drive, Suite 100
Walnut Creek, CA 94598-2540

Date Received: 05/20/14
Work Order: 14-05-1480
Preparation: N/A
Method: EPA 9060A

Project: PALOS VERDE SHELF

Page 1 of 3

Quality Control Sample ID	Type	Matrix	Instrument	Date Prepared	Date Analyzed	LCS/LCSD Batch Number
099-06-013-1057	LCS	Solid	TOC 5	06/06/14	06/06/14 14:40	E0606TOCL1
099-06-013-1057	LCSD	Solid	TOC 5	06/06/14	06/06/14 14:40	E0606TOCL1

Parameter	Spike Added	LCS Conc.	LCS %Rec.	LCSD Conc.	LCSD %Rec.	%Rec. CL	RPD	RPD CL	Qualifiers
Carbon, Total Organic	6000	5442	91	5636	94	80-120	4	0-20	

Return to Contents

RPD: Relative Percent Difference. CL: Control Limits



Calscience

Quality Control - LCS/LCSD

Gilbane
2730 Shadelands Drive, Suite 100
Walnut Creek, CA 94598-2540

Date Received: 05/20/14
Work Order: 14-05-1480
Preparation: EPA 3545
Method: EPA 8270C PEST-SIM

Project: PALOS VERDE SHELF

Page 2 of 3

Quality Control Sample ID	Type	Matrix	Instrument	Date Prepared	Date Analyzed	LCS/LCSD Batch Number
099-13-013-73	LCS	Solid	GC/MS HHH	06/04/14	06/24/14 12:18	140604L31
099-13-013-73	LCSD	Solid	GC/MS HHH	06/04/14	06/24/14 12:54	140604L31

Parameter	Spike Added	LCS Conc.	LCS %Rec.	LCSD Conc.	LCSD %Rec.	%Rec. CL	RPD	RPD CL	Qualifiers
2,4'-DDD	100.0	116.4	116	114.8	115	50-125	1	0-30	
2,4'-DDE	100.0	107.3	107	109.4	109	50-125	2	0-30	
2,4'-DDT	100.0	100.0	100	94.57	95	50-125	6	0-30	
4,4'-DDD	100.0	120.2	120	115.1	115	50-125	4	0-30	
4,4'-DDE	100.0	123.1	123	119.3	119	50-125	3	0-30	
4,4'-DDMU	100.0	96.15	96	94.99	95	50-125	1	0-30	
4,4'-DDNU	100.0	109.3	109	109.4	109	50-125	0	0-30	
4,4'-DDT	100.0	100.9	101	96.86	97	50-125	4	0-30	

Return to Contents

RPD: Relative Percent Difference. CL: Control Limits



Calscience

Quality Control - LCS/LCSD

Gilbane
2730 Shadelands Drive, Suite 100
Walnut Creek, CA 94598-2540

Date Received: 05/20/14
Work Order: 14-05-1480
Preparation: EPA 3545
Method: EPA 8270C SIM PCB Congeners

Project: PALOS VERDE SHELF

Page 3 of 3

Quality Control Sample ID	Type	Matrix	Instrument	Date Prepared	Date Analyzed	LCS/LCSD Batch Number
099-16-418-9	LCS	Solid	GC/MS HHH	06/04/14	06/16/14 21:03	140604L30
099-16-418-9	LCSD	Solid	GC/MS HHH	06/04/14	06/16/14 21:32	140604L30

Parameter	Spike Added	LCS Conc.	LCS %Rec.	LCSD Conc.	LCSD %Rec.	%Rec. CL	ME CL	RPD	RPD CL	Qualifiers
PCB018	50.00	31.56	63	31.27	63	50-150	33-167	1	0-25	
PCB028	50.00	34.25	68	36.12	72	50-150	33-167	5	0-25	
PCB044	50.00	31.59	63	32.32	65	50-150	33-167	2	0-25	
PCB052	50.00	30.25	60	30.79	62	50-150	33-167	2	0-25	
PCB066	50.00	37.05	74	38.45	77	50-150	33-167	4	0-25	
PCB077	50.00	33.66	67	35.48	71	50-150	33-167	5	0-25	
PCB101	50.00	31.56	63	32.15	64	50-150	33-167	2	0-25	
PCB105	50.00	29.47	59	37.34	75	50-150	33-167	24	0-25	
PCB118	50.00	33.60	67	35.05	70	50-150	33-167	4	0-25	
PCB126	50.00	29.30	59	32.82	66	50-150	33-167	11	0-25	
PCB128	50.00	24.95	50	27.78	56	50-150	33-167	11	0-25	
PCB170	50.00	34.81	70	34.11	68	50-150	33-167	2	0-25	
PCB180	50.00	29.08	58	32.79	66	50-150	33-167	12	0-25	
PCB187	50.00	29.19	58	32.99	66	50-150	33-167	12	0-25	
PCB195	50.00	43.44	87	43.43	87	50-150	33-167	0	0-25	
PCB206	50.00	39.37	79	37.71	75	50-150	33-167	4	0-25	
PCB209	50.00	39.27	79	42.40	85	50-150	33-167	8	0-25	

Total number of LCS compounds: 17

Total number of ME compounds: 0

Total number of ME compounds allowed: 1

LCS ME CL validation result: Pass

Return to Contents

RPD: Relative Percent Difference. CL: Control Limits

Glossary of Terms and Qualifiers

Work Order: 14-05-1480

Page 1 of 1

<u>Qualifiers</u>	<u>Definition</u>
*	See applicable analysis comment.
<	Less than the indicated value.
>	Greater than the indicated value.
1	Surrogate compound recovery was out of control due to a required sample dilution. Therefore, the sample data was reported without further clarification.
2	Surrogate compound recovery was out of control due to matrix interference. The associated method blank surrogate spike compound was in control and, therefore, the sample data was reported without further clarification.
3	Recovery of the Matrix Spike (MS) or Matrix Spike Duplicate (MSD) compound was out of control due to suspected matrix interference. The associated LCS recovery was in control.
4	The MS/MSD RPD was out of control due to suspected matrix interference.
5	The PDS/PDSD or PES/PESD associated with this batch of samples was out of control due to suspected matrix interference.
6	Surrogate recovery below the acceptance limit.
7	Surrogate recovery above the acceptance limit.
B	Analyte was present in the associated method blank.
BU	Sample analyzed after holding time expired.
BV	Sample received after holding time expired.
E	Concentration exceeds the calibration range.
ET	Sample was extracted past end of recommended max. holding time.
HD	The chromatographic pattern was inconsistent with the profile of the reference fuel standard.
HDH	The sample chromatographic pattern for TPH matches the chromatographic pattern of the specified standard but heavier hydrocarbons were also present (or detected).
HDL	The sample chromatographic pattern for TPH matches the chromatographic pattern of the specified standard but lighter hydrocarbons were also present (or detected).
J	Analyte was detected at a concentration below the reporting limit and above the laboratory method detection limit. Reported value is estimated.
JA	Analyte positively identified but quantitation is an estimate.
ME	LCS Recovery Percentage is within Marginal Exceedance (ME) Control Limit range (+/- 4 SD from the mean).
ND	Parameter not detected at the indicated reporting limit.
Q	Spike recovery and RPD control limits do not apply resulting from the parameter concentration in the sample exceeding the spike concentration by a factor of four or greater.
SG	The sample extract was subjected to Silica Gel treatment prior to analysis.
X	% Recovery and/or RPD out-of-range.
Z	Analyte presence was not confirmed by second column or GC/MS analysis.
	Solid - Unless otherwise indicated, solid sample data is reported on a wet weight basis, not corrected for % moisture. All QC results are reported on a wet weight basis.

Any parameter identified in 40CFR Part 136.3 Table II that is designated as "analyze immediately" with a holding time of ≤ 15 minutes (40CFR-136.3 Table II, footnote 4), is considered a "field" test and the reported results will be qualified as being received outside of the stated holding time unless received at the laboratory within 15 minutes of the collection time.

A calculated total result (Example: Total Pesticides) is the summation of each component concentration and/or, if "J" flags are reported, estimated concentration. Component concentrations showing not detected (ND) are summed into the calculated total result as zero concentrations.

Recipient Copy

CHAIN-OF-CUSTODY RECORD

Page 16 of 18
COC No. 12926

Order Number: SG012384

Date Shipped: 5/19/2014


AirBill No(s):

From: QATS LABORATORY
2700 CHANDLER AVENUE, BLDG. C
LAS VEGAS, NV 89120
PHONE: 1-702-895-8712
FAX: 1-702-795-8210


To: DANIELLE GONSMAN
CALSCIENCE ENVIRONMENTAL LAB
7440 LINCOLN WAY
GARDEN GROVE CA 92841
714-895-5494

560847650191

14-05-1480

Sample ID	Sigma ID	Qty	Description/Remarks	→ Catalogue Number
SRS1830	SR0668	1	PESTICIDES/CBC-MARINE SEDIMENT	SR-PES
				
SUPERFUND SITE NAME: PALOS VERDE SHELF				

Please use the enclosed Sample Preparation Instructions. If catalogue number(s) are listed at the top of the Sample Preparation Instructions use the Sample Preparation Instructions with catalogue number(s) matching the catalogue number(s) of each of the samples listed above.

Relinquished by: (Signature)	Date/Time 1330 5/19/14	Received by: (Signature)	Date/Time
Custody Seal(s): Present/Absent	Remarks:		
Relinquished by: (Signature)	Date/Time	Received by: (Signature)	Date/Time
			5/20/14 1030

1480

ORIGIN ID: LASA (702) 895-8712
 AIMEE PEREZ
 SHAW ENVIRONMENTAL, INC.
 2700 CHANDLER AVE.
 SUITE C
 LAS VEGAS, NV 89120
 UNITED STATES US

SHIP DATE: 19MAY14
 ACTWGT: 1.0 LB MAN
 CAD: 0157618/CAFE2704



BILL SENDER

TO **ATTN: DANIELLE GONSMAN**
CALSCIENCE ENVIRONMENTAL LAB, INC.
7440 LINCOLN WAY

GARDEN GROVE CA 92841

(714) 896-5494 X 232
 DEPT: KEITH

REF: 03010000



FedEx
 Express




TRK# 5608 4765 0191
 0201

TUE - 20 MAY 10:30A
PRIORITY OVERNIGHT

QZ APVA

92841
 CA-US **SNA**

Part # 156140-434 NRT 6-05



WORK ORDER #: **14-05-** 1 4 8 0

SAMPLE RECEIPT FORM

Box
Cooler 1 of 1

CLIENT: QATS Lab

DATE: 05/20/14

TEMPERATURE: Thermometer ID: SC2 (Criteria: 0.0 °C – 6.0 °C, not frozen except sediment/tissue)

Temperature 20.7 °C - 0.3 °C (CF) = 20.4 °C ☐ Blank ☒ Sample

☒ Sample(s) outside temperature criteria (PM/APM contacted by: LS)

☐ Sample(s) outside temperature criteria but received on ice/chilled on same day of sampling.

☐ Received at ambient temperature, placed on ice for transport by Courier.

Ambient Temperature: ☐ Air ☐ Filter

Checked by: LS

CUSTODY SEALS INTACT:

☐ Cooler ☒ BOX ☐ No (Not Intact) ☐ Not Present ☐ N/A Checked by: LS

☐ Sample ☐ ☐ No (Not Intact) ☒ Not Present Checked by: SR

SAMPLE CONDITION:

	Yes	No	N/A
Chain-Of-Custody (COC) document(s) received with samples.....	☒	☐	☐
COC document(s) received complete.....	☒	☒	☐
☒ Collection date/time, matrix, and/or # of containers logged in based on sample labels.			
☐ No analysis requested. ☐ Not relinquished. ☐ No date/time relinquished.			
Sampler's name indicated on COC.....	☐	☐	☒
Sample container label(s) consistent with COC.....	☒	☐	☐
Sample container(s) intact and good condition.....	☒	☐	☐
Proper containers and sufficient volume for analyses requested.....	☒	☐	☐
Analyses received within holding time.....	☒	☐	☐
Aqueous samples received within 15-minute holding time			
☐ pH ☐ Residual Chlorine ☐ Dissolved Sulfides ☐ Dissolved Oxygen.....	☐	☐	☒
Proper preservation noted on COC or sample container.....	☐	☐	☒
☐ Unpreserved vials received for Volatiles analysis			
Volatile analysis container(s) free of headspace.....	☐	☐	☒
Tedlar bag(s) free of condensation.....	☐	☐	☒

CONTAINER TYPE:

Solid: ☐ 4ozCGJ ☐ 8ozCGJ ☐ 16ozCGJ ☐ Sleeve () ☐ EnCores® ☐ TerraCores® ☒ 2oz PB
 Aqueous: ☐ VOA ☐ VOA_h ☐ VOA_{na2} ☐ 125AGB ☐ 125AGB_h ☐ 125AGB_p ☐ 1AGB ☐ 1AGB_{na2} ☐ 1AGB_s
☐ 500AGB ☐ 500AGJ ☐ 500AGJ_s ☐ 250AGB ☐ 250CGB ☐ 250CGB_s ☐ 1PB ☐ 1PB_{na} ☐ 500PB
☐ 250PB ☐ 250PB_n ☐ 125PB ☐ 125PB_{znna} ☐ 100PJ ☐ 100PJ_{na2} ☐ ☐ ☐

Air: ☐ Tedlar® ☐ Canister Other: ☐ Trip Blank Lot#: Labeled/Checked by: SR

Container: C: Clear A: Amber P: Plastic G: Glass J: Jar B: Bottle Z: Ziploc/Resealable Bag E: Envelope Reviewed by: LS

Preservative: h: HCL n: HNO₃ na₂: Na₂S₂O₃ na: NaOH p: H₃PO₄ s: H₂SO₄ u: Ultra-pure znna: ZnAc₂+NaOH f: Filtered Scanned by: LS

* No Collection date per label